PAS

PREMIER ANALYTICAL SCIENCES

REPORT NO.

PAS-2026-3206
ISSUED 2026-10-03
PAGE 1 OF 1
REVISION 0 — CURRENT

CLIENT

Mohammed

CHAIN

RECEIVED 2026-09-28
CONDITION — INTACT, SEALED
ANALYZED 2026-10-03
RELEASED 2026-10-03 · 11:40 AM PDT
BASIS — PER VIAL

METHODS & LIMITS

a · HPLC-UV / ESI-QTOF-MS — USP/NF 621 Chromatography (SOP-13); LOQ 1 µg/mL
Chromatograms & spectra appended per laboratory number

VERIFY

https://premieranalytical.com/verify/PAS-2026-3206

CERTIFICATE OF ANALYSIS

INDEPENDENT ANALYTICAL LABORATORY · MISSION HILLS, CALIFORNIA

Thymulin

CAS 78922-62-0 · LABEL CLAIM 10 MG · 1 VIAL, NUMBERED · MATERIAL AS STATED BY CLIENT

0 OF 1 SAMPLES PASS Nothing on this certificate was reportable across 1 sample. Each sample below states why.

SAMPLE 01 · LAB NO. PAS-26-00366 · RECEIVED INTACT, SEALED **FLAGGED**

PURITY	% OF LABEL	IDENTITY
—	—	**DOES NOT CONFORM**
Unable to quantify — excessive impurities[a]	Unable to quantify — excessive impurities[a]	HPLC-UV / ESI-QTOF-MS[a]

ALSO MEASURED — Thymulin — Unable to quantify — excessive impurities[a]

PHOTO 1 — AS RECEIVED



PHOTO 2 — LAB NO. AFFIXED

Released by — Michael Khoury, Principal Chemist
released electronically; recorded on the verification register

RELEASED 2026-10-03 · 11:40 AM PDT
REVISION 0 · STATUS: CURRENT

PAS-2026-3206 · SHA-256 0F26 F6A1 D6E7 A180 2FC9 … · VERIFY: https://premieranalytical.com/verify/PAS-2026-3206 · PAGE 1 OF 1

Results relate only to the specific samples tested. Reproduce only in full. This certificate reports only the analyses listed, and its results are intended for research use only — the material tested is not approved for human or veterinary use, or for therapeutic, diagnostic, or clinical applications. Results should be interpreted by qualified professionals within the scope of the intended research. Accuracy and reliability of results may be influenced by sample integrity, handling, and other experimental variables. An endotoxin screen is not a sterility claim. Premier Analytical Sciences does not endorse any company or product listed in this register. A record here states what was measured, nothing more. Premier Analytical Sciences conducts its testing in accordance with ISO/IEC 17025 requirements; it is not an ISO/IEC 17025 accredited laboratory.

/// END OF REPORT ///