

PAS

PREMIER
ANALYTICAL
SCIENCES

REPORT NO.

PAS-2026-3754

ISSUED 2026-10-13

PAGE 1 OF 1

REVISION 0 — CURRENT

CLIENT

Zeus BioLabs

CHAIN

RECEIVED 2026-10-06

CONDITION — INTACT, SEALED

ANALYZED 2026-10-13

RELEASED 2026-10-13 · 10:35 PM

PDT

BASIS — PER VIAL

METHODS & LIMITS

a · Fluorometric — Fluorometric assay
(peptide content) (SOP-13); LOQ 20 µg/mL

b · HPLC-UV / ESI-QTOF-MS — USP/NF
621 Chromatography (SOP-13); LOQ 1
µg/mL

ACCEPTANCE

Purity: 97 % or higher.

Content: 50 to 150 % of the label claim.

Identity: conforms to the reference standard.



VERIFY

[https://premieranalytical.com/verify/
PAS-2026-3754](https://premieranalytical.com/verify/PAS-2026-3754)

CERTIFICATE OF
ANALYSIS

INDEPENDENT ANALYTICAL LABORATORY · MISSION HILLS, CALIFORNIA

Tirzepatide

CAS 2023788-19-2 · LABEL CLAIM 10 MG · CLIENT LOT Z51-Z12 · CLIENT DESIGNATION "GLP-1T" · 1 VIAL,
NUMBERED · MATERIAL AS STATED BY CLIENT

0 OF 1 SAMPLES PASS Purity 94.95 %, content 141.5 % of the stated label claim
across 1 sample, measured per vial by the methods
keyed in the rail.

SAMPLE 01 · LAB NO. PAS-26-00645 · LOT Z51-Z12 · RECEIVED INTACT, SEALED ·
DESIGNATION "GLP-1T"

FLAGGED

Purity 94.95 % is below the 97 % limit.

PURITY

94.95 %

HPLC-UV / ESI-QTOF-MS^b

% OF LABEL

141.5 %

of 10 mg claimed

CONTENT

14.15 mg

Fluorometric^a

IDENTITY

CONFORMS

HPLC-UV / ESI-QTOF-MS^b



PHOTO 1 — AS RECEIVED




PHOTO 2 — LAB NO. AFFIXED

Released by — Michael Khoury, Principal Chemist

released electronically; recorded on the verification register

RELEASED 2026-10-13 · 10:35 PM PDT

REVISION 0 · STATUS: CURRENT

PAS-2026-3754 · SHA-256 FD30 7960 B41A 1FD0 1546 ... · VERIFY: <https://premieranalytical.com/verify/PAS-2026-3754> · PAGE 1 OF 1

Results relate only to the specific samples tested. Reproduce only in full. This certificate reports only the analyses listed, and its results are intended for research use only — the material tested is not approved for human or veterinary use, or for therapeutic, diagnostic, or clinical applications. Results should be interpreted by qualified professionals within the scope of the intended research. Accuracy and reliability of results may be influenced by sample integrity, handling, and other experimental variables. An endotoxin screen is not a sterility claim. Premier Analytical Sciences does not endorse any company or product listed in this register. A record here states what was measured, nothing more. Premier Analytical Sciences conducts its testing in accordance with ISO/IEC 17025 requirements; it is not an ISO/IEC 17025 accredited laboratory.

/// END OF REPORT ///