PAS

PREMIER ANALYTICAL SCIENCES

REPORT NO.

**PAS-2026-4515**
ISSUED 2026-10-09
PAGE 1 OF 1
REVISION 0 — CURRENT

CLIENT

Salus Peptides

CHAIN

RECEIVED 2026-10-02
CONDITION — INTACT, SEALED
ANALYZED 2026-10-09
RELEASED 2026-10-09 · 5:44 PM PDT
BASIS — PER VIAL

METHODS & LIMITS

**a** · Fluorometric — Fluorometric assay (peptide content)
**b** · HPLC-UV / ESI-QTOF-MS — USP/NF 621 Chromatography (SOP-13); LOQ 1 µg/mL
Chromatograms & spectra appended per laboratory number

ACCEPTANCE

Purity: 97 % or higher.
Content: 50 to 150 % of the label claim.
Identity: conforms to the reference standard.

VERIFY

https://premieranalytical.com/verify/PAS-2026-4515

# CERTIFICATE OF ANALYSIS

INDEPENDENT ANALYTICAL LABORATORY · MISSION HILLS, CALIFORNIA

## Ara-290

CAS 1208243-50-8 · LABEL CLAIM 10 MG · CLIENT LOT 26RA101 · CLIENT DESIGNATION "ARA290" · 1 VIAL, NUMBERED · MATERIAL AS STATED BY CLIENT

**0 OF 1 SAMPLES PASS** Purity 82.05 % across 1 sample, measured per vial by the methods keyed in the rail.

SAMPLE 01 · LAB NO. PAS-26-00534 · LOT 26RA101 · RECEIVED INTACT, SEALED · DESIGNATION "ARA290"

**FLAGGED**

Purity 82.05 % is below the 97 % limit.
Ara-290 quantity (content) not reportable: Unable to identify from comparison with standard.

| PURITY | % OF LABEL | IDENTITY |
|---|---|---|
| 82.05 % | — | **CONFORMS** |
| HPLC-UV / ESI-QTOF-MS[b] | Unable to identify from comparison with standard[a] | HPLC-UV / ESI-QTOF-MS[b] |

ALSO MEASURED — Ara-290 — Unable to identify from comparison with standard[a]

PHOTO 1 — AS RECEIVED



PHOTO 2 — LAB NO. AFFIXED

Released by — Michael Khoury, Principal Chemist
released electronically; recorded on the verification register

RELEASED 2026-10-09 · 5:44 PM PDT
REVISION 0 · STATUS: CURRENT

**PAS-2026-4515** · SHA-256 F655 DD0A 7CE6 A365 B84D … · VERIFY: https://premieranalytical.com/verify/PAS-2026-4515 · PAGE 1 OF 1

Results relate only to the specific samples tested. Reproduce only in full. This certificate reports only the analyses listed, and its results are intended for research use only — the material tested is not approved for human or veterinary use, or for therapeutic, diagnostic, or clinical applications. Results should be interpreted by qualified professionals within the scope of the intended research. Accuracy and reliability of results may be influenced by sample integrity, handling, and other experimental variables. An endotoxin screen is not a sterility claim. Premier Analytical Sciences does not endorse any company or product listed in this register. A record here states what was measured, nothing more. Premier Analytical Sciences conducts its testing in accordance with ISO/IEC 17025 requirements; it is not an ISO/IEC 17025 accredited laboratory.

/// END OF REPORT ///