PAS

PREMIER ANALYTICAL SCIENCES

REPORT NO.

PAS-2026-6708

ISSUED 2026-11-09

PAGE 1 OF 1

REVISION 0 — CURRENT

CLIENT

AA immortal labs

CHAIN

RECEIVED 2026-10-27

CONDITION — INTACT, SEALED

ANALYZED 2026-11-05 — 2026-11-09

RELEASED 2026-11-09 · 3:53 PM PST

BASIS — PER VIAL

METHODS & LIMITS

a · HPLC-UV / ESI-QTOF-MS — USP/NF 621 Chromatography (SOP-13); LOQ 1 µg/mL

b · ICP-MS — USP/NF 233 Elemental Impurities—Procedures (limits per 232) LOD / LOQ (µg/g): As 0.033 / 0.099 · Cd 0.028 / 0.085 · Pb 0.049 / 0.147 · Hg 0.015 / 0.045

ACCEPTANCE

Purity: 97 % or higher.

Content: 50 to 150 % of the label claim.

Identity: conforms to the reference standard.

Elemental impurities: each element at or below its limit (ICH Q3D(R2) Option 1, oral).

VERIFY

https://premieranalytical.com/verify/PAS-2026-6708

CERTIFICATE OF ANALYSIS

INDEPENDENT ANALYTICAL LABORATORY · MISSION HILLS, CALIFORNIA

Retatrutide

CAS 2381089-83-2 · LABEL CLAIM 30 MG · CLIENT DESIGNATION "Ratatouille 30mg" · 1 VIAL, NUMBERED · MATERIAL AS STATED BY CLIENT

■ 1 OF 1 SAMPLES PASS Purity 99.05 %, content 141.8 % of the stated label claim across 1 sample, measured per vial by the methods keyed in the rail.

SAMPLE 01 · LAB NO. PAS-26-01148 · RECEIVED INTACT, SEALED · DESIGNATION "Ratatouille 30mg" **■ PASS**

PURITY	% OF LABEL	CONTENT	IDENTITY
99.05 %	141.8 %	42.55 mg	CONFORMS
HPLC-UV / ESI-QTOF-MS[a]	of 30 mg claimed	HPLC-UV / ESI-QTOF-MS[a]	HPLC-UV / ESI-QTOF-MS[a]

ELEMENTAL IMPURITIES[b]

ELEMENT	RESULT	LIMIT	AGAINST LIMIT
Arsenic (As)	Not detected (< 0.033 µg/g)	1.5 µg/g	■ WITHIN LIMIT
Cadmium (Cd)	Not detected (< 0.028 µg/g)	0.5 µg/g	■ WITHIN LIMIT
Lead (Pb)	Not detected (< 0.049 µg/g)	0.5 µg/g	■ WITHIN LIMIT
Mercury (Hg)	Not detected (< 0.015 µg/g)	3 µg/g	■ WITHIN LIMIT



PHOTO 1 — AS RECEIVED




PHOTO 2 — LAB NO. AFFIXED

Released by — Donald Trujillo, Principal Chemist

released electronically; recorded on the verification register

RELEASED 2026-11-09 · 3:53 PM PST

REVISION 0 · STATUS: CURRENT

PAS-2026-6708 · SHA-256 2EED C8EF EAC5 9ED4 D2C2 … · VERIFY: https://premieranalytical.com/verify/PAS-2026-6708 · PAGE 1 OF 1

Results relate only to the specific samples tested. Reproduce only in full. This certificate reports only the analyses listed, and its results are intended for research use only — the material tested is not approved for human or veterinary use, or for therapeutic, diagnostic, or clinical applications. Results should be interpreted by qualified professionals within the scope of the intended research. Accuracy and reliability of results may be influenced by sample integrity, handling, and other experimental variables. An endotoxin screen is not a sterility claim. Premier Analytical Sciences does not endorse any company or product listed in this register. A record here states what was measured, nothing more. Premier Analytical Sciences conducts its testing in accordance with ISO/IEC 17025 requirements; it is not an ISO/IEC 17025 accredited laboratory.

/// END OF REPORT ///